

FirstSave Automated External Defibrillator

Wall Cabinet Procedure



Place key in lock
& turn to set
alarm



When key slot is
horizontal alarm
is set

Remove key



If the centre light is flashing call
LifeTec Medical Limited

Telephone:
0161 941 3597